

EC-Declaration of Conformity on Medical Devices

According to annex II of Medical Device Directive 93/42/EEC (2007/47/EG)

Manufacturer: **RECK-Technik GmbH & Co. KG** Engineering Division Sector
Reckstraße 1–5, 88422 Betzenweiler, GERMANY
phone +49 7374 18-85, fax +49 7374 18-480
email: vigilance@motomed.com

Name of Product: **MOTomed layson edition 261.xxx**
Basic UDI-DI: 42 601 937 10032

| | | REF | SN |
|-------------------|------------------------|---------|--------------|
| Product variants: | MOTomed layson.la | 261.130 | xxx-LY130xxx |
| | MOTomed layson.la | 261.030 | xxx-LY030xxx |
| | MOTomed layson.l | 261.110 | xxx-LY110xxx |
| | MOTomed layson.l | 261.010 | xxx-LY010xxx |
| | MOTomed layson.kidz.la | 261.330 | xxx-LY330xxx |
| | MOTomed layson.kidz.la | 261.230 | xxx-LY230xxx |
| | MOTomed layson.kidz.l | 261.310 | xxx-LY310xxx |
| | MOTomed layson.kidz.l | 261.210 | xxx-LY210xxx |
| | MOTomed layson.l dia | 261.119 | xxx-LY119xxx |
| | MOTomed layson.la prof | 261.139 | xxx-LY139xxx |

Product classification: IIa (rule 9, Medical Device Directive 93/42/EEC)

We hereby declare under our sole responsibility that the above products comply with the relevant provisions of the Medical Devices Directive 93/42/EEC (2007), Annex II, Section 3 and that we are solely responsible for issuing this declaration of conformity.

This Declaration of Conformity is valid until 02.08.2023 or until a revised declaration is made due to product modifications.

Following party was involved in the evaluation of the conformity:
DEKRA Certification GmbH, Handwerkstrasse 15, 70565 Stuttgart, GERMANY
Notified Body No. 0124



Betzenweiler, 20.07.2020

Andreas Reck Executive Director